



National Institute for Public Health
and the Environment
Ministry of Health, Welfare and Sport

Inter-laboratory comparison of particulate matter filter weighing 2022

RIVM report 2022-0117

P. Bongers | P. Panteliadis



National Institute for Public Health
and the Environment
Ministry of Health, Welfare and Sport

Inter-laboratory comparison of particulate matter filter weighing 2022

RIVM report 2022-0117

Colophon

© RIVM 2022

Parts of this publication may be reproduced, provided acknowledgement is given to the: National Institute for Public Health and the Environment, and the title and year of publication are cited.

RIVM attaches a great deal of importance to the accessibility of its products. However, it is at present not yet possible to provide this document in a completely accessible form. If a part is not accessible, it is mentioned as such. Also see www.rivm.nl/en/accessibility

DOI 10.21945/RIVM-2022-0117

P. Bongers (author), RIVM
P. Panteliadis (author), GGD Amsterdam

Contact:
Pleuni Bongers
Centre for Environmental Quality
pleuni.bongers@rivm.nl

This study has been conducted at the own initiative of the participants in this comparison with the aim of checking the quality of reference measurements of particulate matter.

Published by:
**National Institute for Public Health
and the Environment, RIVM**
P.O. Box 1 | 3720 BA Bilthoven
The Netherlands
www.rivm.nl/en

Synopsis

Inter-laboratory comparison of particulate matter filter weighing 2022

Measurements of particulate matter (PM₁₀ and PM_{2,5}) in ambient air are performed by several air monitoring networks in Europe. These PM measurements are carried out by using filters in a monitoring device that ambient air passes through. Every filter is exposed to ambient air for 24 hours. The laboratories of the air monitoring networks calculate the concentration of PM by using the weight of the filters before and after exposure.

In order to investigate whether this weighing procedure – when applied by different laboratories – leads to similar results, a comparison between the laboratories is carried out every two to four years. The weighing results for 2022 were roughly similar.

The results show relatively small differences between the participating laboratories, and these are well within the limit values for PM measurements. This means there is a strong basis for comparing and exchanging PM measurements between the participating laboratories. The filters have to be weighed exactly as described in a European Standard (EN 12341). For example, temperature and relative humidity in the weighing rooms have to fall within specified value ranges. All but one of the participating laboratories matched these requirements.

In 2022, eight laboratories in Europe participated in this research. A total of 84 filters were compared. RIVM performs the measurements together with several air monitoring networks throughout the European Union.

Keywords: particulate matter, PM, reference measurements, filter weighing, inter-laboratory comparison

Publiekssamenvatting

Een vergelijking van de weging van fijnstoffilters door laboratoria in 2022

Verschillende luchtmeetnetten in Nederland en Europa meten de hoeveelheid fijnstof (PM₁₀ en PM_{2,5}) in de buitenlucht. Dit wordt gedaan met filters in een apparaat dat continu buitenlucht aanzuigt. Elk filter wordt na 24 uur vervangen. De laboratoria van de meetnetten wegen het filter voor en na de plaatsing. Hiermee wordt de concentratie fijnstof bepaald.

Elke twee tot vier jaar vergelijkt het RIVM het weegproces bij een aantal van deze laboratoria. Zo wordt gecontroleerd of de fijnstofconcentraties in de verschillende laboratoria overeenkomen. De resultaten van de wegingen waren in 2022 ongeveer gelijkwaardig.

Er waren kleine verschillen te zien die ruim binnen de grenzen vallen die voor fijnstofmetingen zijn bepaald. Dit betekent dat alle fijnstofmetingen onderling goed met elkaar overeenkomen en tussen de deelnemende meetnetten kunnen worden uitgewisseld. De laboratoria moeten de metingen volgens een Europees vastgestelde procedure doen (EN 12341). De weegkamers horen bijvoorbeeld een bepaalde temperatuur en luchtvochtigheid te hebben. Op één na voldeden alle deelnemende laboratoria aan deze procedure.

In 2022 deden acht laboratoria uit Europa mee aan de vergelijking. In totaal zijn 84 filters gemeten. Het RIVM doet de metingen samen met Nederlandse meetnetten en enkele lidstaten van de Europese Unie. Het RIVM is het nationaal referentielaboratorium voor luchtkwaliteitsmetingen in de buitenlucht en heeft daarom deze taak.

Kernwoorden: fijnstof, PM, referentie-metingen, filterweging, inter-laboratorium vergelijking

Contents

Summary – 9

1 Introduction – 11

2 Methods – 13

3 Results and evaluations – 15

3.1 Results – 15

3.1.1 General – 15

3.1.2 Results of the pilot laboratory – 15

3.1.3 Results of the participants – 16

3.2 Evaluation of the weighing results – 16

3.2.1 Laboratory performance – 16

3.2.2 Method performance – 18

4 Conclusions – 21

Annex 1 Filters supplied to each laboratory – 25

Annex 2 Results of weighings in this study – 28

Annex 3 Weighing room conditions of every participant – 30

Annex 4 Laboratory performance evaluation – 33

Summary

Reference measurements of particulate matter (PM) in ambient air are performed by sampling an accurately known volume of air through a filter for 24 hours. By measuring the mass difference of the filter before and after sampling, the concentration of particulate matter can be determined. For this purpose, filters are weighed at least twice, both before and after sampling, under strict conditions of temperature and relative humidity, as described in European Standard EN 12341:2014.

In order to investigate whether application of this procedure – when applied by different laboratories – leads to comparable results, an inter-laboratory comparison has been conducted. For this purpose, one pilot laboratory has sent sets of twelve pre-weighed filters typically used for low-volume sampling (eight loaded, four blanks) to other participating laboratories. After having been weighed by participants, the filters have been reweighed by the pilot laboratory.

In addition, the temperatures and relative humidity values in the participants' weighing rooms have been monitored using portable temperature/humidity meters. The monitoring results indicate that one laboratory has problems in controlling the temperature and relative humidity in their weighing room.

Evaluation of the results of the comparison shows that – when average results of the pilot laboratory are used as assigned reference values and En numbers are calculated using estimated weighing uncertainties – all results meet the criterion for En to be ≤ 1 .

From the results, the inter-laboratory (reproducibility) uncertainty of the weighing results has been calculated and has been found to be 0.057 mg for a confidence level of 95% for all filters. This corresponds to a relative uncertainty of 2.1% at the level of daily limit value for PM₁₀ (50 µg/m³) and 3.4% at the level of a substitute daily limit value for PM_{2,5} (30 µg/m³).

1 Introduction

Determination of filter masses is an essential part of the application of the European reference methods for the measurement of PM₁₀ and PM_{2,5} concentrations in ambient air. The procedures for filter weighing are described in European Standard EN 12341:2014 [1]. They comprise subsequent conditioning and weighing under prescribed temperature and relative humidity conditions at the weighing facilities. Such conditions are essential in order to avoid significant changes in filter masses due to (de)sorption of water vapour – both by filter materials and particulate matter – and evaporation of semi-volatile constituents of particulate matter.

In order to investigate the comparability of the mass determination of both loaded and blank filters, an inter-laboratory comparison has been conducted in February to April, 2022. Participating laboratories, including their abbreviations and locations, are (in alphabetical order):

- DCMR Rijnmond Environmental Service (DCMR; Schiedam, the Netherlands);
- Environmental Research Laboratory of the National Centre for Scientific Research 'Demokritos' (EREL; Athens, Greece);
- Flanders Environment Agency (VMM; Antwerp, Belgium);
- National Institute for Public Health and the Environment (RIVM; Bilthoven, the Netherlands);
- Public Health Service of Amsterdam (GGD Amsterdam; Amsterdam, the Netherlands), pilot laboratory;
- Scientific Institute of Public Service (ISSeP; Liège, Belgium);
- South Limburg Regional Implementing Agency (RUDZL; Maastricht, the Netherlands);
- State Agency for Nature, Environment and Consumer Protection (LANUV; Essen, Germany).

This report describes the used methods, results, and findings of this comparison.

2 Methods

The comparison is based on the weighing of filters typically used for low-volume sampling in accordance with EN 12341:2014 [1], with a diameter of 47 mm. 56 loaded filters and 28 blank filters have been collected by the pilot laboratory (GGD Amsterdam) from the following sources:

- 28 Whatman QMA quartz-fibre filters from rural monitoring sites (RIVM);
- 14 Whatman QMA quartz-fibre filters from urban background and traffic sites (RIVM);
- 14 Whatman QMA quartz-fibre filters from traffic sites (RIVM);
- 28 pre-conditioned blank Whatman QMA quartz-fibre filters (RIVM).

All filters have been conditioned in the pilot laboratory's weighing facilities for a minimum of 48 hours at (20 ± 1) °C and $(47.5 \pm 2.5)\%$ relative humidity (% RH). After conditioning, the filters have been subjected to four consecutive weighings, with a minimum interval of 24 hours between weighings (from February 21st to March 1st, 2022). When the results of the first and fourth weighings differ by less than 0.060 mg for loaded filters and 0.040 mg for blank filters, the mean results for all filters are used as the 'mass before despatch'.

Subsequently, from the set of filters, seven subsets of twelve filters have been prepared, each containing eight loaded filters and four blank filters from the sources mentioned above, and despatched to the other participants.

All participants have stored, conditioned, and weighed the filters in their own facilities according to the procedure described in [1] as follows; conditioning for a minimum of 48 hours followed by a first weighing, conditioning again for a minimum of 24 hours followed by a second weighing. If the two results differ by more than 0.040 mg for blank filters and 0.060 mg for sampled filters, the procedure for the second weighing is repeated for a third weighing and the average of the last two consecutive weighings are taken as the filter mass. If these requirements aren't met, the filters are not taken into account in this study.

In addition, all participants have been supplied with portable temperature/relative humidity sensors to log readings every minute. Filters and sensors have been transported to the participants in cool boxes. The sensors have been used to record temperatures and relative humidity values in the participants' weighing facilities over a period enclosing the actual conditioning and weighing periods.

After their return, filters and sensors have been placed in the pilot laboratory's weighing room. Filters have been reconditioned and reweighed four times following the prescribed conditions (from March 30th to April 11th, 2022). When the results of the first and fourth weighings differ by less than 0.060 mg for loaded filters and 0.040 mg

for blank filters, the mean results for all filters are used as the 'mass after return'.

In Annex 1, the codes and (types of) PM loadings of the filters provided to the participants are given.

The sensors have been re-calibrated in the pilot laboratory's weighing room by comparison of their readings with those of a calibrated dew-point meter. The results of the calibrations have been used to correct the readings.

3 Results and evaluations

3.1 Results

3.1.1 *General*

The results of the weighings are presented in Annex 2 as follows:

- The left-hand column contains filter codes;
- the second and third columns contain the weighing results of the pilot laboratory before and after despatch to participants;
- the fourth column provides the mean result that is used as the reference value;
- the fifth and sixth columns contain the two weighing results of each participant;
- the seventh column present the third weighing result of each participant in case this was needed;
- the eighth column contains the mean participant results;
- the ninth column shows the difference between reference values and participants' mean results.

3.1.2 *Results of the pilot laboratory*

As a criterion for calculating the mean mass before despatch for every filter, the first and fourth weighings had to differ by less than 0.060 mg for loaded filters and 0.040 mg for blank filters. For all filters, this criterion has been fulfilled. The same criterion applied to calculating the mean mass after return, for which all filters matched this criterion.

From the results of the weighings performed by the pilot laboratory before and after despatch, it is observed that, on average, the mass after return of the loaded filters is lower than it was before despatch (mean difference of -0.086 mg), whereas the mass for blank filters is on average almost the same (mean difference of -0.002 mg). The last (fourth) weighing of loaded filters before despatch also showed a (minor) decrease of mass, indicating that the filters were already losing volatiles during the five days they were stored in the pilot laboratory.

The mass changes of the loaded filters are plotted against the masses of PM originally determined on the filters in Figure 1. The relation observed between the mass change and the filter loading suggests some proportionality between the mass loss and the mass of PM on the filter. This may be an indication for losses of (semi-)volatile constituents of the PM.




Figure 1 Mass changes of sampled filters after return to the pilot laboratory.

3.1.3 Results of the participants

Differences between mean participants' results and reference values range from -0.014 mg to +0.124 mg, with an average of 0.021 mg.

Differences between two consecutive weighings are all within the requirements of EN 12341:2014 (≤ 0.040 mg for blank filters, ≤ 0.060 mg for sampled filters).

Corrected results of sensor measurements of temperatures and relative humidity values in weighing facilities of participants (one-hour average values) are presented graphically in Annex 3. They reveal that for the weighing facilities of most participating laboratories, the temperature is stable and relative humidity shows a small upward trend during the placement period of some sensors. Overall – with the exception of laboratory 2 – the mean average weighing room temperature and relative humidity values are within the requirements given in [1]. It is observed that measurements in the weighing room of laboratory 2 exhibit fluctuations and are on average below the minimum temperature and minimum relative humidity requirements.

Reports by participants confirm this trend and show that all weighings for this study – except for those in laboratory 2 – have (within the inherent uncertainties of the loggers) been performed at 45-50% RH and 19-21 °C.

3.2 Evaluation of the weighing results

3.2.1 Laboratory performance

For the evaluation of the results of the comparison, the En number has been used. En is calculated as follows [2]:

$$En = \frac{|x_p - x_{ref}|}{2 \cdot \sqrt{u_p^2 + u_{ref}^2}}$$

where

x_p is the participant's result;
 x_{ref} is the reference value;
 u_p is the measurement uncertainty of the participant's result;
 u_{ref} is the measurement uncertainty of the reference value.

An En number ≤ 1 is an indication of comparability of the participant's result and the reference value at a level of confidence of 95%.

The mean results of consecutive weighings of each filter performed by the participants have been taken as values of x_p ; the mean results of the weighings before and after performed by the pilot laboratory have been taken as reference values.

The measurement uncertainty of the values of x_p has been estimated according to the approach described in EN 12341:2014 as follows:

$$u_{x,p}^2 = u_{cal}^2 + u_{zd}^2 + u_{buoy}^2 + u_{\Delta x,p}^2$$

where

u_{cal} is the uncertainty in the calibration of the balance used for the weighings;
 u_{zd} is the uncertainty due to zero drift of the balance;
 u_{buoy} is the uncertainty due to differences in buoyancy;
 $u_{\Delta x,p}$ is the uncertainty due to the difference between the two weighing results of the participants.

The first three uncertainty contributions have been calculated from the maximum criteria and default values given in EN 12341:2014, resulting in a value of 0.0084 mg. The uncertainty due to the difference between the two weighing results of the participant has been calculated by assuming a uniform distribution as:

$$u_{\Delta x,p}^2 = \frac{(x_{p,2} - x_{p,1})^2}{12}$$

where $x_{p,1}$ and $x_{p,2}$ are the participant's two weighing results.

The uncertainty of the reference values is estimated similarly as:

$$u_{x,ref}^2 = u_{cal}^2 + u_{zd}^2 + u_{buoy}^2 + u_{\Delta x,ref}^2$$

where $u_{\Delta x,ref}$ is the uncertainty due to the difference between the results of the pilot laboratory before and after despatch of the filters.

The uncertainty due to the difference between the two weighing results of the pilot laboratory has been calculated by assuming a uniform distribution as:

$$u_{\Delta x, ref}^2 = \frac{(x_{ref,2} - x_{ref,1})^2}{12}$$

where $x_{ref,1}$ and $x_{ref,2}$ are the weighing results before despatch and after return of the filters.

Because of potential differences in weighing room relative humidity values between participants and pilot laboratory, an additional (default) uncertainty contribution has been added based on the maximum criteria for blank and loaded filter masses included in EN 12341:2014:

- 0.060/ $\sqrt{3}$ mg for loaded filters;
- 0.040/ $\sqrt{3}$ mg for blank filters.

The results of all calculations are given in Annex 4. These show that for all results, En numbers are ≤ 1 , indicating correspondence with the reference values at 95% probability.

3.2.2 Method performance

3.2.2.1 Uncertainty from differences

The uncertainty of the weighing method may be estimated from the differences of the results of participants and the pilot laboratory as follows:

$$u^2 = \frac{\sum_{i=1}^n (x_{i,p} - x_{i,ref})^2}{2n}$$

where

- $x_{i,p}$ is the participant's result for sample i ;
- $x_{i,ref}$ is the reference value for sample i ;
- n is the number of samples.

The resulting uncertainty is 0.023 mg.

When splitting the results between sampled and blank filters only, the uncertainty is 0.027 mg for sampled filters and 0.011 mg for blank filters.

Combining both uncertainties in quadrature may yield an indication of the uncertainty of a differential weighing. This results in an uncertainty of 0.029 mg and an expanded uncertainty (95% confidence) of 0.058 mg.

3.2.2.2 ISO 5725 part 2 approach

The availability of normalized measurement results permits the evaluation of the method performance by application of the statistics of

ISO 5725 part 2 [3] by treating normalized results for one laboratory as replicate results. Normalized results are calculated as follows:

$$x_{i,n} = \frac{x_{i,p}}{x_{i,ref}}$$

where

$x_{i,p}$ is the participant's result for sample i ;
 $x_{i,ref}$ is the reference value for sample i .

Calculation of ISO 5725 part 2 statistical parameters for results of this study yields the relative standard deviations shown in Table 1 – with a differentiation between results for loaded filters and blank filters, and all filters combined. Using the mean mass of all filters, the absolute statistics may be estimated by multiplying the relative standard deviations by these mean masses. The results show that the weighing method used is quite robust.

Table 1 Relative standard deviations (RSD) and absolute standard deviations (SD) for loaded and blank filters and all filters combined.

	Loaded filters	Blank filters	All filters
Relative standard deviations			
Repeatability RSD	0.00008	0.00004	0.00006
Between-laboratory RSD	0.00017	0.00009	0.00014
Reproducibility RSD	0.00019	0.00009	0.00016
Absolute standard deviations			
Repeatability SD (mg)	0.0114	0.0058	0.0095
Between-laboratory SD (mg)	0.026	0.013	0.021
Reproducibility SD (mg)	0.029	0.014	0.024

The relative standard deviations are comparable to those found in the previous comparison in 2018 [4].

In addition, the results are submitted to Mandel's h and k tests for outliers of mean values and standard deviations, respectively, and to Grubb's and Cochran's outlier tests.

A closer examination of the results of the ISO 5725 part 2 evaluation reveals that:

- For all normalized data, one result of Laboratory 1, three results of Laboratory 2 and six results of Laboratory 4 are marked as outliers of mean (99% probability), as indicated by Mandel's h statistics. These are indicated with a superscript h in Annex 2.
- For all data pooled, five results of Laboratory 2 and two results of Laboratory 7 are marked as outliers of precision (99% probability), as indicated by Mandel's k statistics, marked with a superscript k in Annex 4.
- One of the results of Laboratory 2 is marked as an outlier of precision (99% probability), as indicated by Cochran's test statistics, marked with an asterisk in Annex 4.

- No outliers of mean (99% probability) were found by conducting Grubb's tests for the single largest and smallest observations.

As some weighings from Laboratory 2 are marked as outliers in multiple test statistics and because all of these weighings have not been performed at 19-21 °C and most not at 45-50% RH, they will be discarded. The individual outliers from Laboratory 1, 4, and 7 could be explained by the fact that the results are obtained from filters from different sources and therefore they don't necessarily have the same PM composition. Probably a loss of volatiles or humidity was taking place during the five days of storage of the filters in the pilot laboratory, resulting in higher absolute losses for filters with higher PM masses. When omitting all the results of Laboratory 2 from the evaluation, the (relative) standard deviations as shown in Table 2 are obtained.

Table 2 Relative standard deviations (RSD) and absolute standard deviations (SD) for loaded and blank filters and all filters combined, after removal of outliers.

	Loaded filters	Blank filters	All filters
Relative standard deviations			
Repeatability RSD	0.00005	0.00004	0.00004
Between-laboratory RSD	0.00017	0.00009	0.00014
Reproducibility RSD	0.00018	0.00010	0.00015
Absolute standard deviations			
Repeatability SD (mg)	0.0069	0.0053	0.0064
Between-laboratory SD (mg)	0.025	0.013	0.021
Reproducibility SD (mg)	0.026	0.014	0.022

When using the reproducibility standard deviations after removal of the outlying results as measures of the uncertainties of weighing results of loaded and blank filters, the method uncertainty may be calculated by combining both results in quadrature:

$$u_m^2 = s_{R,l}^2 + s_{R,b}^2$$

where $s_{R,l}$ and $s_{R,b}$ are the reproducibility standard deviations for loaded and blank filter weighings, respectively.

The combination results in an uncertainty value of 0.029 mg and an expanded uncertainty (95% confidence) of 0.057 mg. This value is similar to the value obtained from differences in results between participants and pilot laboratory (0.058 mg).

4 Conclusions

Organization

The comparison involves the transport of filters to and from participating laboratories with filter weighings performed by the pilot laboratory before and after both transportations. It is observed that during this process the mass of loaded filters decreases on average, potentially due to losses of (semi)volatile constituents of particulate matter. Also, the fact that the sampled Whatman QMA filters have been preconditioned at high-relative humidity values may make losses of water an obvious cause of mass decrease.

The mass decreases are incorporated into the uncertainty budget used for the evaluation of laboratory performances.

Laboratory performance

Overall, the results of the laboratory performance evaluation are satisfactory. Evaluation using En numbers shows that all laboratories, even with outliers, meet the performance requirement of $En \leq 1$.

A study of the control of temperatures and relative humidity values in the weighing facilities of participants shows that all except one of the laboratories confirm with the requirements in European Standard EN 12341:2014 [1].

Method performance

The method performance, expressed as the reproducibility standard deviation of weighing results calculated according to ISO 5725 part 2 without outliers, is 0.022 mg. When results of loaded filters only are used, the reproducibility standard deviation increases slightly to 0.026 mg.

By combining uncertainties for the weighing of loaded and blank filters expressed as reproducibility standard deviations, an uncertainty of 0.029 mg results, leading to an expanded method uncertainty of 0.057 mg. When using this value to calculate the relative uncertainty contribution of the weighing process to the overall uncertainty of the measurement of PM, expressed at the limit values of PM₁₀ and PM_{2,5}, the following values are obtained (95% confidence):

- PM₁₀ at 50 µg/m³ (daily limit value [5]): 2.1%;
- PM₁₀ at 45 µg/m³ (daily air quality guideline level of the World Health Organization (WHO) [6]): 2.3%;
- PM_{2,5} at 30 µg/m³ (substitute daily limit value [7]): 3.4%;
- PM_{2,5} at 15 µg/m³ (daily guideline level of the WHO [6]): 6.9%.

These values suggest that the uncertainty contributions of the filter conditioning and weighing process as obtained in this study are relatively small, considering that the maximum allowed uncertainty for PM reference measurements is 25% [5].

References

- [1] European Committee for Standardization. EN 12341:2014. Ambient air – Standard gravimetric measurement method for the determination of the PM₁₀ or PM_{2,5} mass concentration of suspended particulate matter. Brussels: CEN.
- [2] International Organization for Standardization. ISO 13528:2015(Cor. 2016-10). Statistical methods for use in proficiency testing by interlaboratory comparison. Geneva: ISO.
- [3] International Organization for Standardization. ISO 5725-2:2019(E). Accuracy (trueness and precision) of measurement methods and results – Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method. Geneva: ISO.
- [4] GGD Amsterdam (2018) Inter-laboratory comparison filter weighing 2018. Doc 18-1164. Amsterdam: GGD Amsterdam.
- [5] European Parliament, Council of the European Union. Directive 2008/50/EC of the European Parliament and of the Council of 21 May 2008 on ambient air quality and cleaner air for Europe, Annex I – Annex XI. Official Journal of the European Union, L 152.
- [6] World Health Organization (2021) WHO global air quality guidelines. Particulate matter (PM_{2.5} and PM₁₀), ozone, nitrogen dioxide, sulfur dioxide and carbon monoxide, p. 74-97. Geneva: World Health Organization.
- [7] European Committee for Standardization. EN 16450:2017. Ambient air – Automated measuring systems for the measurement of the concentration of particulate matter (PM₁₀; PM_{2,5}). Brussels: CEN.

Annex 1 Filters supplied to each laboratory

Laboratory 1 VMM

Filter Code	PM type	PM mass (mg)
A1	PM ₁₀	1.028
A2	PM ₁₀	1.587
A3	PM ₁₀	1.020
A4	PM ₁₀	1.041
A5	PM ₁₀	0.548
A6	PM ₁₀	0.974
A7	PM _{2,5}	0.550
A8	PM _{2,5}	0.377
B1	blank	-
B2	blank	-
B3	blank	-
B4	blank	-

Laboratory 2 EREL

Filter Code	PM type	PM mass (mg)
A9	PM ₁₀	1.306
A10	PM _{2,5}	0.643
A11	PM ₁₀	0.947
A12	PM _{2,5}	1.084
A13	PM ₁₀	0.777
A14	PM _{2,5}	0.403
A15	PM _{2,5}	0.587
A16	PM ₁₀	1.030
B5	blank	-
B6	blank	-
B7	blank	-
B8	blank	-

Laboratory 3 ISSeP

Filter Code	PM type	PM mass (mg)
A17	PM _{2,5}	0.985
A18	PM _{2,5}	0.201
A19	PM ₁₀	1.713
A20	PM _{2,5}	0.389
A21	PM _{2,5}	0.496
A22	PM ₁₀	0.833
A23	PM ₁₀	0.972
A24	PM _{2,5}	0.722
B9	blank	-
B10	blank	-
B11	blank	-
B12	blank	-

Laboratory 4 LANUV

Filter Code	PM type	PM mass (mg)
A25	PM ₁₀	1.293
A26	PM _{2,5}	0.224
A27	PM _{2,5}	0.744
A28	PM _{2,5}	0.586
A29	PM ₁₀	1.506
A30	PM ₁₀	1.551
A31	PM _{2,5}	1.050
A32	PM _{2,5}	0.855
B13	blank	-
B14	blank	-
B15	blank	-
B16	blank	-

Laboratory 5 DCMR

Filter Code	PM type	PM mass (mg)
A33	PM _{2,5}	1.269
A34	PM ₁₀	1.538
A35	PM _{2,5}	0.105
A36	PM ₁₀	1.751
A37	PM ₁₀	1.104
A38	PM _{2,5}	0.453
A39	PM ₁₀	1.123
A40	PM ₁₀	1.211
B17	blank	-
B18	blank	-
B19	blank	-
B20	blank	-

Laboratory 6 RUDZL

Filter Code	PM type	PM mass (mg)
A41	PM _{2,5}	1.172
A42	PM ₁₀	0.818
A43	PM _{2,5}	0.866
A44	PM ₁₀	1.684
A45	PM ₁₀	1.467
A46	PM ₁₀	1.709
A47	PM _{2,5}	0.934
A48	PM _{2,5}	1.277
B21	blank	-
B22	blank	-
B23	blank	-
B24	blank	-

Laboratory 7 RIVM

Filter Code	PM type	PM mass (mg)
A49	PM ₁₀	0.945
A50	PM _{2,5}	0.629
A51	PM ₁₀	1.366
A52	PM _{2,5}	0.657
A53	PM _{2,5}	0.416
A54	PM _{2,5}	0.623
A55	PM _{2,5}	0.787
A56	PM ₁₀	1.227
B25	blank	-
B26	blank	-
B27	blank	-
B28	blank	-

Annex 2 Results of weighings in this study

Filter Code	Pilot			Participant				Difference
	Before	After	Mean	First	Second	Third	Mean	
A1	149.782	149.668	149.725	149.717	149.704		149.710	0.014
A2	148.184	148.057	148.120	148.129 ^h	148.114		148.122	-0.002
A3	145.381	145.366	145.374	145.378	145.369		145.373	0.000
A4	149.184	149.072	149.128	149.131	149.120		149.126	0.002
A5	147.051	147.025	147.038	147.025	147.009		147.017	0.021
A6	151.046	150.933	150.989	150.989	150.975		150.982	0.007
A7	147.158	147.086	147.122	147.118	147.106		147.112	0.010
A8	148.816	148.780	148.798	148.807	148.787		148.797	0.000
B1	146.416	146.423	146.419	146.421	146.421		146.421	-0.002
B2	147.529	147.537	147.533	147.547	147.546		147.547	-0.014
B3	143.326	143.339	143.332	143.333	143.328		143.331	0.002
B4	148.302	148.286	148.294	148.298	148.296		148.297	-0.003
A9	150.211	150.099	150.155	150.035	150.085	150.040 ^h	150.063	0.092
A10	147.996	147.931	147.963	147.915 ^h	147.945		147.930	0.033
A11	147.253	147.162	147.207	147.145	147.140		147.143	0.065
A12	149.121	149.033	149.077	149.030	149.070		149.050	0.027
A13	144.100	144.051	144.076	144.050	144.035		144.043	0.033
A14	148.156	148.082	148.119	148.125	148.090		148.108	0.011
A15	147.122	147.051	147.086	147.110	147.070		147.090	-0.004
A16	150.245	150.119	150.182	150.165	150.090 ^h	150.150	150.120	0.062
B5	149.142	149.127	149.134	149.135	149.115		149.125	0.009
B6	146.382	146.373	146.377	146.355	146.350		146.353	0.025
B7	149.277	149.265	149.271	149.255	149.245		149.250	0.021
B8	148.111	148.092	148.102	148.100	148.100		148.100	0.001
A17	147.899	147.770	147.834	147.817	147.813		147.815	0.019
A18	147.375	147.373	147.374	147.362	147.364		147.363	0.011
A19	146.840	146.694	146.767	146.746	146.742		146.744	0.023
A20	149.957	149.883	149.920	149.895	149.897		149.896	0.024
A21	149.008	148.940	148.974	148.957	148.958		148.958	0.017
A22	149.039	148.941	148.990	148.987	148.992		148.990	0.001
A23	146.915	146.790	146.853	146.831	146.831		146.831	0.022
A24	146.790	146.664	146.727	146.700	146.701		146.701	0.026
B9	146.802	146.812	146.807	146.794	146.799		146.797	0.010
B10	147.599	147.603	147.601	147.594	147.596		147.595	0.006
B11	149.112	149.111	149.111	149.120	149.121		149.121	-0.009
B12	148.164	148.145	148.154	148.153	148.157		148.155	-0.001


All weighing results are in milligrams. The difference is calculated by subtracting the mean weighing result of the participant laboratory by the mean weighing result of the pilot laboratory.

^h Outlier as indicated by Mandel's *h* statistics (99% probability).


Filter Code	Pilot			Participant				Difference
	Before	After	Mean	First	Second	Third	Mean	
A25	149.664	149.498	149.581	149.505	149.507		149.506	0.075
A26	147.851	147.832	147.841	147.826	147.821		147.824	0.018
A27	149.548	149.422	149.485	149.434	149.431		149.433	0.052
A28	147.379	147.276	147.327	147.277	147.272		147.275	0.053
A29	149.864	149.664	149.764	149.657 ^h	149.643 ^h		149.650	0.113
A30	151.737	151.558	151.647	151.529 ^h	151.517 ^h		151.523	0.124
A31	148.076	147.936	148.006	147.933	147.924 ^h		147.929	0.078
A32	144.282	144.181	144.231	144.169	144.158		144.164	0.068
B13	148.295	148.278	148.286	148.267	148.268		148.268	0.019
B14	144.626	144.610	144.618	144.594	144.600		144.597	0.021
B15	147.488	147.469	147.478	147.453	147.460		147.457	0.022
B16	141.993	141.966	141.979	141.942 ^h	141.950		141.946	0.033
A33	144.825	144.791	144.808	144.801	144.800		144.801	0.007
A34	147.106	147.037	147.072	147.052	147.049		147.051	0.021
A35	148.723	148.721	148.722	148.719	148.719		148.719	0.003
A36	150.056	149.969	150.012	149.991	149.988		149.990	0.023
A37	146.882	146.883	146.882	146.877	146.875		146.876	0.006
A38	148.876	148.856	148.866	148.860	148.859		148.860	0.006
A39	149.565	149.510	149.537	149.522	149.522		149.522	0.015
A40	149.116	149.055	149.085	149.064	149.078		149.071	0.014
B17	146.539	146.563	146.551	146.550	146.552		146.551	0.000
B18	147.421	147.435	147.428	147.424	147.425		147.425	0.003
B19	148.455	148.471	148.463	148.457	148.459		148.458	0.005
B20	146.379	146.400	146.389	146.386	146.387		146.387	0.002
A41	146.840	146.737	146.788	146.764	146.765		146.765	0.024
A42	146.519	146.445	146.482	146.461	146.462		146.462	0.020
A43	147.177	147.080	147.128	147.119	147.123		147.121	0.007
A44	145.420	145.307	145.363	145.344	145.354		145.349	0.014
A45	149.243	149.114	149.178	149.148	149.142		149.145	0.033
A46	148.624	148.482	148.553	148.531	148.528		148.530	0.024
A47	147.051	146.973	147.012	147.008	147.012		147.010	0.002
A48	148.938	148.846	148.892	148.890	148.889		148.890	0.002
B21	146.460	146.470	146.465	146.463	146.469		146.466	-0.001
B22	148.331	148.269	148.300	148.267	148.275		148.271	0.029
B23	147.402	147.428	147.415	147.419	147.429		147.424	-0.009
B24	146.905	146.913	146.909	146.913	146.924		146.919	-0.010
A49	148.303	148.235	148.269	148.245	148.254		148.250	0.019
A50	148.671	148.595	148.633	148.599	148.616		148.608	0.025
A51	146.500	146.392	146.446	146.419	146.439		146.429	0.017
A52	147.577	147.512	147.544	147.517	147.536		147.527	0.018
A53	147.054	147.006	147.030	147.003	147.016		147.010	0.021
A54	148.771	148.701	148.736	148.709	148.720		148.715	0.021
A55	148.517	148.447	148.482	148.453	148.466		148.460	0.023
A56	150.442	150.370	150.406	150.372	150.396		150.384	0.022
B25	146.752	146.761	146.756	146.730	146.743		146.737	0.020
B26	146.806	146.805	146.805	146.778	146.791		146.785	0.021
B27	148.003	147.998	148.000	147.976	147.984		147.980	0.020
B28	146.451	146.454	146.453	146.423	146.445		146.434	0.019

Annex 3 Weighing room conditions of every participant

Temperature logs from the weighing facility of VMM




Relative humidity logs from the weighing facility of VMM




Weighing room conditions of laboratory 1

Temperature logs from the weighing facility of EREL




Relative humidity logs from the weighing facility of EREL




Weighing room conditions of laboratory 2

Temperature logs from the weighing facility of ISSeP




Relative humidity logs from the weighing facility of ISSeP




Weighing room conditions of laboratory 3

Temperature logs from the weighing facility of LANUV




Relative humidity logs from the weighing facility of LANUV




Weighing room conditions of laboratory 4

Temperature logs from the weighing facility of DCMR




Relative humidity logs from the weighing facility of DCMR




Weighing room conditions of laboratory 5

Temperature logs from the weighing facility of RUDZL




Relative humidity logs from the weighing facility of RUDZL




Weighing room conditions of laboratory 6

Temperature logs from the weighing facility of RIVM



Relative humidity logs from the weighing facility of RIVM



Weighing room conditions of laboratory 7

Annex 4 Laboratory performance evaluation

Filter code	Pilot		Participant		En number
	Mean	Uncertainty	Mean	Uncertainty	
A1	149.725	0.034	149.710	0.009	0.15
A2	148.120	0.038	148.122	0.009	0.02
A3	145.374	0.009	145.373	0.009	0.00
A4	149.128	0.033	149.126	0.009	0.02
A5	147.038	0.011	147.017	0.009	0.28
A6	150.989	0.034	150.982	0.009	0.07
A7	147.122	0.023	147.112	0.009	0.12
A8	148.798	0.013	148.797	0.010	0.00
B1	146.419	0.009	146.421	0.008	0.03
B2	147.533	0.009	147.547	0.008	0.27
B3	143.332	0.009	143.331	0.008	0.03
B4	148.294	0.009	148.297	0.008	0.05
A9	150.155	0.034	150.063 ^k	0.015	0.91
A10	147.963	0.020	147.930	0.012	0.40
A11	147.207	0.028	147.143	0.008	0.72
A12	149.077	0.027	149.050 ^{k,*}	0.014	0.29
A13	144.076	0.016	144.043	0.009	0.42
A14	148.119	0.023	148.108 ^k	0.013	0.13
A15	147.086	0.022	147.090 ^k	0.014	0.04
A16	150.182	0.037	150.120 ^k	0.019	0.57
B5	149.134	0.009	149.125	0.010	0.17
B6	146.377	0.009	146.353	0.008	0.47
B7	149.271	0.009	149.250	0.009	0.40
B8	148.102	0.010	148.100	0.008	0.03
A17	147.834	0.038	147.815	0.008	0.18
A18	147.374	0.008	147.363	0.008	0.15
A19	146.767	0.043	146.744	0.008	0.20
A20	149.920	0.023	149.896	0.008	0.29
A21	148.974	0.021	148.958	0.008	0.20
A22	148.990	0.029	148.990	0.008	0.01
A23	146.853	0.037	146.831	0.008	0.21
A24	146.727	0.037	146.701	0.008	0.26
B9	146.807	0.009	146.797	0.008	0.20
B10	147.601	0.008	147.595	0.008	0.12
B11	149.111	0.008	149.121	0.008	0.18
B12	148.154	0.010	148.155	0.008	0.01

Weighing results (in mg), uncertainties (in mg) and En numbers.

^k Outlier as indicated by Mandel's *k* statistics (99% probability).

* Outlier as indicated by Cochran's test statistics (99% probability).

Filter code	Pilot		Participant		En number
	Mean	Uncertainty	Mean	Uncertainty	
A25	149.581	0.048	149.506	0.008	0.62
A26	147.841	0.010	147.824	0.008	0.24
A27	149.485	0.037	149.433	0.008	0.51
A28	147.327	0.031	147.275	0.008	0.56
A29	149.764	0.058	149.650	0.009	0.83
A30	151.647	0.052	151.523	0.009	0.98
A31	148.006	0.041	147.929	0.009	0.71
A32	144.231	0.030	144.164	0.009	0.72
B13	148.286	0.010	148.268	0.008	0.36
B14	144.618	0.010	144.597	0.008	0.39
B15	147.478	0.010	147.457	0.009	0.41
B16	141.979	0.011	141.946	0.009	0.61
A33	144.808	0.013	144.801	0.008	0.10
A34	147.072	0.022	147.051	0.008	0.25
A35	148.722	0.008	148.719	0.008	0.03
A36	150.012	0.027	149.990	0.008	0.26
A37	146.882	0.008	146.876	0.008	0.09
A38	148.866	0.010	148.860	0.008	0.09
A39	149.537	0.018	149.522	0.008	0.19
A40	149.085	0.020	149.071	0.009	0.18
B17	146.551	0.011	146.551	0.008	0.00
B18	147.428	0.009	147.425	0.008	0.06
B19	148.463	0.010	148.458	0.008	0.09
B20	146.389	0.010	146.387	0.008	0.05
A41	146.788	0.031	146.765	0.008	0.25
A42	146.482	0.023	146.462	0.008	0.24
A43	147.128	0.029	147.121	0.008	0.08
A44	145.363	0.034	145.349	0.009	0.14
A45	149.178	0.038	149.145	0.008	0.32
A46	148.553	0.042	148.530	0.008	0.22
A47	147.012	0.024	147.010	0.008	0.02
A48	148.892	0.028	148.890	0.008	0.03
B21	146.465	0.009	146.466	0.008	0.03
B22	148.300	0.020	148.271	0.009	0.46
B23	147.415	0.011	147.424	0.009	0.17
B24	146.909	0.009	146.919	0.009	0.18
A49	148.269	0.021	148.250	0.009	0.23
A50	148.633	0.024	148.608	0.010	0.30
A51	146.446	0.032	146.429 ^k	0.010	0.17
A52	147.544	0.020	147.527	0.010	0.22
A53	147.030	0.016	147.010	0.009	0.26
A54	148.736	0.022	148.715	0.009	0.26
A55	148.482	0.022	148.460	0.009	0.27
A56	150.406	0.022	150.384	0.011	0.25
B25	146.756	0.009	146.737	0.009	0.38
B26	146.805	0.008	146.785	0.009	0.39
B27	148.000	0.008	147.980	0.009	0.39
B28	146.453	0.008	146.434 ^k	0.010	0.35

RIVM

Committed to *health and sustainability*